



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, BATHINDA

ਅਖਿਲ ਭਾਰਤੀ ਆਯੁਰ ਵਿਗਿਆਨ ਸੰਸਥਾਨ, ਬਠਿੰਡਾ ਅਖਿਲ ਭਾਰਤੀ ਆਯੁਰਵਿਗਿਆਨ ਸੰਸਥਾਨ, ਬਠਿੰਡਾ

Procurement Cell



CORRIGENDUM

1. In reference to NIQ AIIMS Bathinda/Proc.Cell/2025/201 dated 26 Sep 2025, published on the AIIMS Bathinda website with the due date 07 Oct 2025, find the revised specification at Annexure – A.
2. Rest no change.

for
30.09.2025
OIC, Procurement Cell

Specification For Lab Refrigerator

- Capacity: >900 ltrs
- Temperature Range From +2°C To 8°C or adjustable
- Set point at 4°C
- Accuracy $\pm 1^\circ\text{C}$
- Atleast 10 adjustable shelves
- Exterior made up of High quality galvanized metal sheets.
- Double layered toughened glass door
- Air forced cooling
- CFC free Refrigerant
- Hermetically sealed compressor
- Microprocessor controlled
- Door type: Glass door
- Controller cum display and sensor Failure alarm
- Visual and acoustic alarms like adjustable High/Low temperature alarm, power failure alarm, open door alarm.
- System must be provided with Data logger for maintaining temperature log
- Certified.


DR. ABHINAV KANWAL
Associate Professor
Department of Pharmacology
AHS Bhinda

Aman

Narayan