



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, BATHINDA
Mandi Dabwali Road, Bathinda (Pb.) – 151001

PROCUREMENT CELL

AIIMS Bathinda/Proc.Cell/2026/219

Start Date: 13/07/2026
Due Date: 20/07/2026

Notice Inviting Quotations (NIQ)

The following items are required for AIIMS, Mandi Dabwali Road, Bathinda – 151001:

Sr.	Product with Description	Qty.	Specifications
I.	Automated External Defibrillators (AEDs) for Mobile Units (MMUs)	02	As per Annexure - A

Kindly arrange to send your quotation giving the lowest rates per unit along with terms and conditions in a sealed cover addressed to the **Executive Director, AIIMS Bathinda, Mandi Dabwali Road, Bathinda – 151001, Punjab**, to reach this office before the last date of receiving Quotations, i.e., on **20/07/2026, Time 17:00**

The word “Quotations for Automated External Defibrillators (AEDs) for Mobile Units (MMUs) for the Dept. of CFM at AIIMS Bathinda, our reference number, and the date should boldly be mentioned on the cover.

- While submitting the quotation following should invariably be mentioned:
 - Details of Specification and its lowest rate F.O.R. AIIMS Bathinda.
 - GST Registration Number (attach copy)
 - PAN No. (Attach copy)
 - Clearly mentioned GST rates.
 - Tax at a concessional rate as applicable to the Educational Institutions
 - Completion from the date of receipt of the confirmed order.
- The rates will be valid for one year from the date of issuance of the Work Order.
- Payment will be made after the supply and inspection of the delivered items.
- Under no circumstances will unsealed quotations be entertained in the office.
- Quotations received after the due date shall not be considered.


11/7/26
for OIC, Procurement Cell

Quotation Proforma
(To be returned with Quotation)

Following proforma should be filled in and duly signed by the firm and sent along with the quotation.

Firm's Reference			Date				
Firm Registration No. (if any)			PAN				
GST No.		Address:					
Phone No.							
Email Id							
M/s:			NIQ No.				
			Last Date of Submission				
Sr.	Product with Description	Qty	Unit Price	GST	Total Unit Price	Total Value	Total Value in words
1							
2							
3							
Total							

- I/we engage to supply the material(s) as required to your office and comply the following:
- i. Terms and conditions of the NIQ.
 - ii. That the offer is valid for 90 days.
 - iii. That we have not been debarred by any Government body/PSU/Autonomous Body
 - iv. That the rates quoted are not higher than the rates quoted for same item to any Government body/PSU/Autonomous Body.
 - v. That the quotation submitted by us is properly sealed and prepared so as to prevent any subsequent alteration and replacement.

Place:
Date:

Authorized Signatory
(seal)

Terms and Conditions

1. The quotation must be in the form furnished by Procuring Entity and should be free from corrections/erasures. In case there is any unavoidable correction it should be properly attested. If not, the quotation will not be considered. Quotation written in pencil will not be considered.
2. All India Institute of Medical Sciences (AIIMS), Bathinda reserves the right to accept the offer by individual items and reject any or all quotations without assigning any reason thereof and does not bind itself to accept lowest quotations.
3. Manufacturer's name and country of origin of materials offered must be clearly specified. Please quote whether your organisation is large scale industry or small scale industry. If you have NSIC/MSE/MSI/DGS&D Certificate, please attach it to the quotation. Mention your registration details.
4. Complete details and ISI specification if any must accompany the quotation. Make/brand of the item shall be stated wherever applicable. If you have got any counter offer as suitable to the material required by us, the same may be shown separately.
5. Samples must be submitted where specified along with the quotations. Samples must be carefully packed, sealed and labelled clearly with enquiry number, subject and sender's name for easy identification. Rejected samples will be returned at your cost if insisted.
6. All supplies are subject to inspection and approval before acceptance. Manufacturer/supplier warranty certificates and manufacturer/Government approved lab test certificate shall be furnished along with the supply, wherever applicable.
7. AIIMS Bathinda reserves the right to modify the quantity specified in this enquiry.
8. The prices quoted should be firm till the supplies are completed. Please quote the rates in words and figures. Rates quoted should be free delivery at destination including all charges otherwise the quotation is likely to be rejected. Prices quoted for free delivery at destination will be given preference. If there is no indication regarding the FOR, in the quotation, then it will be considered as FOR destinations. Price quoted should be net and valid for a minimum period of 90 days from the date of opening of the quotation.
9. Delivery period required for supplying the material should be invariably specified in the quotation.
10. In case your quotation is accepted and order is placed on you, the supply against the order should be made within the period stipulated in the order. AIIMS Bathinda reserves the right to recover any loss sustained due to delayed delivery by way of penalty. Failure to supply the material within the stipulated period shall entitle Procuring Entity for the imposition of penalty without assigning any reasons @ 0.5% (half percent) of the total value of the item covered in order as penalty per day subject to a maximum of 5% (five percent) unless extension is obtained in writing from the office on valid ground before expiry of delivery period.
11. If the deliveries are not maintained and due to that account Procuring Entity is forced to buy the material at your risk and cost from elsewhere, the loss or damage that may be sustained there by will be recovered from the defaulting supplier.
12. Dispute clause: Any dispute relating to the enquiry shall be subject to the jurisdiction of the court at Bathinda only.

Generic Technical Specifications: Automated External Defibrillator (AED)

- AED should be Semi automatic, quick and simple to use by anyone in a hospital or in a public place. It should analyze ECG, Charge and deliver the shock all by itself without operator's intervention.
- It should have an easy operation process in 3 steps, START-ANALYSIS-SHOCK.
- AED should be based on Biphasic waveform Technology and should have facility of escalating energy level from 150 to 360Joules for Adults and 50 Joules for Child.
- Selection of Child Mode by connecting Child Pad to machine.
- Shock delivery via adhesive pads
- It should provide interval in between shocks for performing CPR as per latest AHA guidelines.
- AED should have upgradable for connecting a CPR sensor in future whenever required.
- Shock to shock cycle time should be less than 25 seconds between shocks in series.
- Maximum Charging Time: 50J < 2 Seconds; 150J < 5 Seconds; 200J < 6 Seconds; 270J < 8 Seconds; 360J < 10 Seconds
- The AED must have a Status Indicator on front panel.
- AED machine should guide user through both audible and visual text display means of instruction.
- The Battery should have capacity of more than 300 shocks at 200 Joules or 15 hours of continuous monitoring.
- Environmental operation condition: Temperature between 0 - 50C and humidity maximum to 95%.
- AED should have water and dust Ingress protection : IP56
- The Machine should do the automatic self-test when switched ON.
- AED must be capable of storing 100 events or 2 hours of ECG recording and ambient sound recording up to 10 hours.
- Battery must be having Shelf life of 5 Years.
- AED Electrode Pads should have Shelf life of 02 years
- AED Must have Language selection possibilities. it should work on International English and Hindi Language.
- AED must be capable of transferring saved events and recording to a PC/Laptop .
- Charging time is less than 6 seconds to charge to maximum energy of 360 Joules.
- Weight should not be more than 2kg including battery.
- AED Device must be CDSCO Registered and valid certificate should be available.
- AED should be USFDA or European CE Certificate issued by notified body .
- AED should have IEC 60601-2-4 Certificate .
- AED Must supply with standard Accessories :
- A Pair of Disposable Adult size Adhesive Pads – 1 No
- AED preparation Kit containing 1 Pair of surgical gloves, 1 scissor and 1 CPR mask – 1 Set
- Transport Bag – 1 No
- USB Cable – 1 No
- Instruction manual and PC software CD.